

August 31, 2018

Rob King
Hampton Bays Water District
P.O. Box 1013
Hampton Bays, NY 11946

RE: Project: HOMEOWNER BOTTLED WATER 8/24
Pace Project No.: 7063157

Dear Rob King:

Enclosed are the analytical results for sample(s) received by the laboratory on August 29, 2018. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,



Stu Murrell
stu.murrell@pacelabs.com
(631)694-3040
Project Manager

Enclosures

cc: Warren Booth, Hampton Bays Water District
John Collins, H2M Group
Stella Michaels, Hampton Bays Water District
Paul Ponturo, H2M Group



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

Long Island Certification IDs

575 Broad Hollow Rd, Melville, NY 11747

New York Certification #: 10478 Primary Accrediting Body

New Jersey Certification #: NY158

Pennsylvania Certification #: 68-00350

Connecticut Certification #: PH-0435

Maryland Certification #: 208

Rhode Island Certification #: LAO00340

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

Lab ID	Sample ID	Matrix	Date Collected	Date Received
7063157001	NESTLE PURE BOTTLED WATER	Drinking Water	08/24/18 07:30	08/29/18 16:30

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

Lab ID	Sample ID	Method	Analysts	Analytes Reported
7063157001	NESTLE PURE BOTTLED WATER	EPA 200.7	JMW	7
		EPA 200.8	SK2	12
		SM22 4500 NH3 H	BNK	1

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

Sample: NESTLE PURE BOTTLED WATER **Lab ID: 7063157001** Collected: 08/24/18 07:30 Received: 08/29/18 16:30 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	Reg. Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.7 MET ICP, Drinking Water		Analytical Method: EPA 200.7							
Calcium	11.4	mg/L	0.20		1		08/31/18 14:36	7440-70-2	
Ca Hardness as CaCO ₃ (SM 2340B)	28.5	mg/L	0.50		1		08/31/18 14:36		
Iron	0.056	mg/L	0.020		1		08/31/18 14:36	7439-89-6	
Magnesium	4.2	mg/L	0.20		1		08/31/18 14:36	7439-95-4	
Manganese	0.034	mg/L	0.010		1		08/31/18 14:36	7439-96-5	
Sodium	8.3	mg/L	0.20		1		08/31/18 14:36	7440-23-5	
Tot Hardness asCaCO ₃ (SM 2340B)	45.8	mg/L	0.83		1		08/31/18 14:36		N3
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8							
Antimony	0.68	ug/L	0.40		1		08/31/18 12:51	7440-36-0	
Arsenic	<1.0	ug/L	1.0		1		08/31/18 12:51	7440-38-2	
Barium	<0.0020	mg/L	0.0020		1		08/31/18 12:51	7440-39-3	
Beryllium	<0.30	ug/L	0.30		1		08/31/18 12:51	7440-41-7	
Cadmium	<1.0	ug/L	1.0		1		08/31/18 12:51	7440-43-9	
Chromium	<0.0070	mg/L	0.0070		1		08/31/18 12:51	7440-47-3	
Lead	<1.0	ug/L	1.0		1		08/31/18 12:51	7439-92-1	
Mercury	<0.20	ug/L	0.20		1		08/31/18 12:51	7439-97-6	
Nickel	<0.00050	mg/L	0.00050		1		08/31/18 12:51	7440-02-0	
Selenium	<2.0	ug/L	2.0		1		08/31/18 12:51	7782-49-2	
Silver	<0.0010	mg/L	0.0010		1		08/31/18 12:51	7440-22-4	
Thallium	<0.30	ug/L	0.30		1		08/31/18 12:51	7440-28-0	
4500 Ammonia Water		Analytical Method: SM22 4500 NH ₃ H							
Nitrogen, Ammonia	0.14	mg/L	0.10		1		08/30/18 13:42	7664-41-7	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

QC Batch: 81390	Analysis Method: EPA 200.7
QC Batch Method: EPA 200.7	Analysis Description: 200.7 MET No Prep Drinking Water
Associated Lab Samples: 7063157001	

METHOD BLANK: 374392 Matrix: Drinking Water

Associated Lab Samples: 7063157001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L	<0.50	0.50	08/31/18 14:26	
Calcium	mg/L	<0.20	0.20	08/31/18 14:26	
Iron	mg/L	<0.020	0.020	08/31/18 14:26	
Magnesium	mg/L	<0.20	0.20	08/31/18 14:26	
Manganese	mg/L	<0.010	0.010	08/31/18 14:26	
Sodium	mg/L	<0.20	0.20	08/31/18 14:26	
Tot Hardness asCaCO3 (SM 2340B)	mg/L	<0.83	0.83	08/31/18 14:26	N3

LABORATORY CONTROL SAMPLE: 374393

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L		60.9			
Calcium	mg/L	25	24.4	98	85-115	
Iron	mg/L	2	2.0	99	85-115	
Magnesium	mg/L	25	24.8	99	85-115	
Manganese	mg/L	.25	0.25	99	85-115	
Sodium	mg/L	50	48.8	98	85-115	
Tot Hardness asCaCO3 (SM 2340B)	mg/L		163			N3

MATRIX SPIKE SAMPLE: 374396

Parameter	Units	7063223001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L	153000 ug/L		214			
Calcium	mg/L	61300 ug/L	25	85.7	98	70-130	
Iron	mg/L	123 ug/L	2	2.2	103	70-130	
Magnesium	mg/L	4020 ug/L	25	29.6	102	70-130	
Manganese	mg/L	<10.0 ug/L	.25	0.26	104	70-130	
Sodium	mg/L	35000 ug/L	50	85.2	100	70-130	
Tot Hardness asCaCO3 (SM 2340B)	mg/L	170000 ug/L		336			N3

MATRIX SPIKE SAMPLE: 374398

Parameter	Units	7063223002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L	360000 ug/L		415			

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

MATRIX SPIKE SAMPLE: 374398

Parameter	Units	7063223002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Calcium	mg/L	144000 ug/L	25	166	88	70-130	
Iron	mg/L	141 ug/L	2	2.2	105	70-130	
Magnesium	mg/L	603 ug/L	25	26.5	104	70-130	
Manganese	mg/L	<10.0 ug/L	.25	0.26	104	70-130	
Sodium	mg/L	34800 ug/L	50	84.9	100	70-130	
Tot Hardness asCaCO3 (SM 2340B)	mg/L	362000 ug/L		524			N3

SAMPLE DUPLICATE: 374395

Parameter	Units	7063223001 Result	Dup Result	RPD	Max RPD	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L	153000 ug/L	150	2	20	
Calcium	mg/L	61300 ug/L	60.1	2	20	
Iron	mg/L	123 ug/L	0.12	2	20	
Magnesium	mg/L	4020 ug/L	3.9	3	20	
Manganese	mg/L	<10.0 ug/L	<0.010		20	
Sodium	mg/L	35000 ug/L	34.3	2	20	
Tot Hardness asCaCO3 (SM 2340B)	mg/L	170000 ug/L	166	2	20	N3

SAMPLE DUPLICATE: 374397

Parameter	Units	7063223002 Result	Dup Result	RPD	Max RPD	Qualifiers
Ca Hardness as CaCO3 (SM 2340B)	mg/L	360000 ug/L	357	1	20	
Calcium	mg/L	144000 ug/L	143	1	20	
Iron	mg/L	141 ug/L	0.15	6	20	
Magnesium	mg/L	603 ug/L	0.60	0	20	
Manganese	mg/L	<10.0 ug/L	<0.010		20	
Sodium	mg/L	34800 ug/L	34.5	1	20	
Tot Hardness asCaCO3 (SM 2340B)	mg/L	362000 ug/L	360	1	20	N3

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

QC Batch: 81365	Analysis Method: EPA 200.8
QC Batch Method: EPA 200.8	Analysis Description: 200.8 MET No Prep Drinking Water
Associated Lab Samples: 7063157001	

METHOD BLANK: 374311 Matrix: Water

Associated Lab Samples: 7063157001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Antimony	ug/L	<0.40	0.40	08/31/18 13:03	
Arsenic	ug/L	<1.0	1.0	08/31/18 13:03	
Barium	mg/L	<0.0020	0.0020	08/31/18 13:03	
Beryllium	ug/L	<0.30	0.30	08/31/18 13:03	
Cadmium	ug/L	<1.0	1.0	08/31/18 13:03	
Chromium	mg/L	<0.0070	0.0070	08/31/18 13:03	
Lead	ug/L	<1.0	1.0	08/31/18 13:03	
Mercury	ug/L	<0.20	0.20	08/31/18 13:03	
Nickel	mg/L	<0.00050	0.00050	08/31/18 13:03	
Selenium	ug/L	<2.0	2.0	08/31/18 13:03	
Silver	mg/L	<0.0010	0.0010	08/31/18 13:03	
Thallium	ug/L	<0.30	0.30	08/31/18 13:03	

LABORATORY CONTROL SAMPLE: 374312

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	50	46.6	93	85-115	
Arsenic	ug/L	50	47.5	95	85-115	
Barium	mg/L	.05	0.048	96	85-115	
Beryllium	ug/L	50	48.1	96	85-115	
Cadmium	ug/L	50	47.9	96	85-115	
Chromium	mg/L	.05	0.048	96	85-115	
Lead	ug/L	50	47.8	96	85-115	
Mercury	ug/L	1.5	1.4	92	85-115	
Nickel	mg/L	.05	0.047	94	85-115	
Selenium	ug/L	50	47.3	95	85-115	
Silver	mg/L	.05	0.048	97	85-115	
Thallium	ug/L	50	48.9	98	85-115	

MATRIX SPIKE SAMPLE: 374315

Parameter	Units	7063223001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Antimony	ug/L	0.75	10	11.0	102	70-130	
Arsenic	ug/L	<1.0	4	4.8	112	70-130	
Barium	mg/L	90.8 ug/L	.2	0.30	107	70-130	
Beryllium	ug/L	<0.30	5	4.6	91	70-130	
Cadmium	ug/L	<1.0	5	5.5	109	70-130	
Chromium	mg/L	<7.0 ug/L	.02	0.024	102	70-130	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

MATRIX SPIKE SAMPLE: 374315		7063223001	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Lead	ug/L	173	2	169	-164	70-130	M1
Mercury	ug/L	<0.20	.4	0.26	55	70-130	M1
Nickel	mg/L	113 ug/L	.05	0.16	100	70-130	
Selenium	ug/L	<2.0	1	2.1	90	70-130	
Silver	mg/L	<1.0 ug/L	.005	0.0033	64	70-130	M1
Thallium	ug/L	<0.30	5	5.2	104	70-130	

MATRIX SPIKE SAMPLE: 374418		7063304001	Spike	MS	MS	% Rec	
Parameter	Units	Result	Conc.	Result	% Rec	Limits	Qualifiers
Antimony	ug/L	<0.40	10	10.4	104	70-130	
Arsenic	ug/L	<1.0	4	4.8	111	70-130	
Barium	mg/L	11.9 ug/L	.2	0.23	111	70-130	
Beryllium	ug/L	<0.30	5	5.7	115	70-130	
Cadmium	ug/L	<1.0	5	5.7	114	70-130	
Chromium	mg/L	<7.0 ug/L	.02	0.022	107	70-130	
Lead	ug/L	3.0	2	5.0	103	70-130	
Mercury	ug/L	<0.20	.4	0.35	86	70-130	
Nickel	mg/L	144 ug/L	.05	0.20	104	70-130	
Selenium	ug/L	<2.0	1	<2.0	115	70-130	
Silver	mg/L	<1.0 ug/L	.005	0.0044	87	70-130	
Thallium	ug/L	<0.30	5	5.2	103	70-130	

SAMPLE DUPLICATE: 374314		7063223001	Dup	RPD	Max	
Parameter	Units	Result	Result	RPD	RPD	Qualifiers
Antimony	ug/L	0.75	0.74	1	20	
Arsenic	ug/L	<1.0	<1.0		20	
Barium	mg/L	90.8 ug/L	0.090	1	20	
Beryllium	ug/L	<0.30	<0.30		20	
Cadmium	ug/L	<1.0	<1.0		20	
Chromium	mg/L	<7.0 ug/L	<0.0070		20	
Lead	ug/L	173	172	0	20	
Mercury	ug/L	<0.20	<0.20		20	
Nickel	mg/L	113 ug/L	0.11	0	20	
Selenium	ug/L	<2.0	<2.0		20	
Silver	mg/L	<1.0 ug/L	<0.0010		20	
Thallium	ug/L	<0.30	<0.30		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

SAMPLE DUPLICATE: 374417

Parameter	Units	7063304001 Result	Dup Result	RPD	Max RPD	Qualifiers
Antimony	ug/L	<0.40	<0.40		20	
Arsenic	ug/L	<1.0	<1.0		20	
Barium	mg/L	11.9 ug/L	0.012	1	20	
Beryllium	ug/L	<0.30	<0.30		20	
Cadmium	ug/L	<1.0	<1.0		20	
Chromium	mg/L	<7.0 ug/L	<0.0070		20	
Lead	ug/L	3.0	3.0	3	20	
Mercury	ug/L	<0.20	<0.20		20	
Nickel	mg/L	144 ug/L	0.14	0	20	
Selenium	ug/L	<2.0	<2.0		20	
Silver	mg/L	<1.0 ug/L	<0.0010		20	
Thallium	ug/L	<0.30	<0.30		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

QC Batch: 81169	Analysis Method: SM22 4500 NH3 H
QC Batch Method: SM22 4500 NH3 H	Analysis Description: 4500 Ammonia
Associated Lab Samples: 7063157001	

METHOD BLANK: 373284 Matrix: Water
Associated Lab Samples: 7063157001

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Nitrogen, Ammonia	mg/L	<0.10	0.10	08/30/18 12:49	

LABORATORY CONTROL SAMPLE: 373285

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	1	1.1	109	90-110	

MATRIX SPIKE SAMPLE: 373286

Parameter	Units	7062506001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Nitrogen, Ammonia	mg/L	19.7	10	29.9	102	75-125	

SAMPLE DUPLICATE: 373287

Parameter	Units	7062506001 Result	Dup Result	RPD	Max RPD	Qualifiers
Nitrogen, Ammonia	mg/L	19.7	18.0	9	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

M1 Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

N3 Accreditation is not offered by the relevant laboratory accrediting body for this parameter.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: HOMEOWNER BOTTLED WATER 8/24

Pace Project No.: 7063157

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
7063157001	NESTLE PURE BOTTLED WATER	EPA 200.7	81390		
7063157001	NESTLE PURE BOTTLED WATER	EPA 200.8	81365		
7063157001	NESTLE PURE BOTTLED WATER	SM22 4500 NH3 H	81169		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

WO#: 7063157



7063157
1747
5

Sample Request Form PUBLIC WATER SUPPLIER

Date: 8-24-18

8/24/18

1320

WELL OFF LINE

WELL RUN TO SYSTEM

Client Info:

Name or Code: HAMPTON BAYS WATER DISTRICT

P.O. BOX 1013

Address: HAMPTON BAYS, NEW YORK 11946

(631) 728-0179

Phone #: _____

Attn: _____

Proj. # or (Name): _____

Bill To: _____

Copies To: _____

Collected By: Home Officer
Accepted By: [Signature]
Cooler Temp: 5.5 °C

YES NO VOC'S PRESERVED WITH HCl

Back at 1030

Sample Types	Purpose	Origin	Treatment Types
PW - Potable Water	RO - Routine	D - Distribution	AST - Air Stripper
GW - Groundwater	RE - Resample	RW - Raw Well	GAC - Granular Activated Charcoal
SW - Surface Water	S - Special	TW - Treated Well	N - Nitrate Removal Plant
WW - Waste Water		T - Tank	FE - Iron Removal Plant
AQ - Aqueous		MW - Monitoring Well	O - Other
S - Soil		I - Influent	
		E - Effluent	

Sample Info:

Date/Time Collected:	Sample Type	Location	Origin	Treatment Type	Purpose	Field Readings Cl ₂	pH/Temp	Analysis	Lab No.
8-24-18 7:30	PW	Homeowner Bottled water Nestle Pure Life	-	-	S			Pretals / Ammonia	

Remarks:



Sample Condition Upon Receipt

Client Name: HBW

Project **WO#: 7063157**
 PM: SWM Due Date: 09/07/18
 CLIENT: HBW

Courier: Fed Ex UPS USPS Client Commercial Pace Other

Tracking #: _____
 Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No

Packing Material: Bubble Wrap Bubble Bags Ziploc None Other

Thermometer Used: TH091 Correction Factor: 0.0 Samples on ice, cooling process has begun

Cooler Temperature (°C): 3.3 Cooler Temperature Corrected (°C): 3.3 Date/Time 5035A kits placed in freezer _____

Temp should be above freezing to 6.0°C

USDA Regulated Soil (N/A, water sample) Date and Initials of person examining contents: [Signature] 9/29/18

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? YES NO
 Did samples originate from a foreign source (internationally, including Hawaii and Puerto Rico)? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (F-LI-C-010) and include with SCUR/COC paperwork.

		COMMENTS:
Chain of Custody Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr):	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	6.
Rush Turn Around Time Requested:	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note if sediment is visible in the dissolved container.
Sample Labels match COC:	<input type="checkbox"/> Yes <input type="checkbox"/> No	12.
-Includes date/time/ID/Analysis Matrix SL WT OIL		
All containers needing preservation have been checked	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # <u>HL739245</u>		Sample #
All containers needing preservation are found to be in compliance with EPA recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, NaOH > 12 Cyanide)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	Initial when completed: Lot # of added preservative: Date/Time preservative added
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/BO15 (water). Per Method, VOA pH is checked after analysis		
Samples checked for dechlorination:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot #		
Residual chlorine strips Lot #		
Headspace in VOA Vials (>6mm):	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15.
Trip Blank Present:	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Trip Blank Custody Seals Present	<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Pace Trip Blank Lot # (if applicable): _____		

Client Notification/ Resolution: _____ Field Data Required? Y / N
 Person Contacted: _____ Date/Time: _____
 Comments/ Resolution: _____

